


Sermorelin HPLC Test Result

Date: 11/01/2019

一、Test Method:

Column: :C18, 4.6*250mm, 5um
Solvent B :100% Acetonitrile
Solvent A :0.1% Trifluoroacetic in 100% Water
Gradient : B A
0.0min 30% 70%
30.0min 60% 40%
Flow Rate :1.0ml/min
Wavelength :220nm
Temperature :37°C
Concentration :0.5mg/ml
Sample Volume :20ul

二、Auto-Scaled Chromatogram:



三、Result:

	RT [min]	ST [min]	ET [min]	Area [mAU.s]	Area [%]
1	9.705	9.352	9.806	59.208	0.1
2	10.012	9.806	10.124	173.186	0.4
3	10.217	10.124	10.358	165.982	0.4
4	11.115	10.768	12.798	39663.231	98.9
5	12.866	12.798	13.023	19.435	0.0
6	13.380	13.260	13.535	40.476	0.1
	Total			40121.516	100.0