


TB500 HPLC Test Result

Date: 11/01/2019

一、 Test Method:

Column: :C18, 4.6*250mm, 5um
Solvent B :100% Acetonitrile
Solvent A :0.1% Trifluoroacetic in 100% Water
Gradient : B A
0.0min 10% 90%
30.0min 40% 60%
Flow Rate :1.0ml/min
Wavelength :220nm
Temperature :37°C
Concentration :1.0mg/ml
Sample Volume :20ul

二、 Auto-Scaled Chromatogram:



三、 Result:

	RT [min]	ST [min]	ET [min]	Area [mAU.s]	Area [%]
1	17.209	17.055	17.442	44.129	0.6
2	17.858	17.684	17.971	66.497	0.9
3	18.302	17.971	19.591	6934.067	98.4
	合计			7044.692	100.0