

HPLC REPORT

Sample Description:

Structure : TB500

Cas No : 77591-33-4

Sequence : Ac-Ser-Asp-Lys-Pro-Asp-Met-Ala-Glu-Ile-Glu-Lys-Phe-Asp-Lys-Ser-Lys-Leu-Lys-Lys-Thr-Glu-Thr-Gln-Glu-Lys-Asn-Pro-Leu-Pro-Ser-Lys-Glu-Thr-Ile-Glu-Gln-Glu-Lys-Gln-Ala-Gly-Glu-Ser-OH

Number : 010250011

Lot No : P191120-YS055820

Column : 4.6×250mm, Venusil XBP-C18

Solvent A : 0.1% trifluoroacetic in 100% acetonitrile


Solvent B : 0.1% trifluoroacetic in 100% water

Gradient	:	A	B
	0.01min	23%	77%
	25min	48%	52%
	25.1min	100%	0%
	30min	STOP	

Flow rate : 1.0 mL/min

Wavelength : 220nm

Volume : 5ul



Peak No.	Ret Time	Height	Area	Conc.
1	10.648	665.293	9599.668	0.4646
2	11.095	2101.169	21470.002	1.0390
3	11.367	195403.594	2031587.250	98.3171
4	12.048	355.692	3705.901	0.1793
Total				100.0000