


# HPLC TEST RESULT

## CJC-1295 DAC

Structure: CJC-1295 DAC  
 Cas No :  
 Sequence: Tyr-D-Ala-Asp-Ala-Ile-Phe-Thr-Gln-Ser-Tyr-Arg-Lys-Val-Leu-Ala-Gln-Leu-Ser-Ala-Arg-Lys-Leu-Leu-Gln-Asp-Ile-Leu-Ser-Arg-Lys  
 Lot No: P191025-LR087292  
 Number: 0200049  
 Column: 250\*4.6mm, Kromasil-C18-5um  
 Solvent A: 0.1%TFA in 100%water  
 Solvent B: 0.1%TFA in 100%acetonitrile  
 Gradient:
 

	A	B
0.1min	67%	33%
25.0min	56%	44%
25.1min	0%	100%
30.0min	stop	

  
 Flow Rate: 1.0ml/min  
 Wavelength(nm): 220  
 Volume: 10ul



Peak No	Ret Time	Height	Area	Conc.
1	11.650	1386.122	19445.871	0.3768
2	11.937	3810.708	16066.445	0.3113
3	12.148	451912.250	4908521.000	95.1126
4	12.148	13324.289	216712.938	4.1993

100.00

# HPLC TEST RESULT

## CJC-1295 DAC

### REPORT

Sequence :

