


# HPLC TEST RESULT

## CJC1295

Structure: CJC1295  
 Cas No : 863288-34-0  
 Sequence: Tyr-D-Ala-Asp-Ala-Ile-Phe-Thr-Gln-Ser-Tyr-Arg-Lys-Val-Leu-Ala-Gln-Leu-Ser-Ala-Arg-Lys-Leu-Leu-Gln-Asp-Ile-Leu-Ser-Arg-Lys-NH<sub>2</sub>  
 Lot No: P191029-LR088421  
 Number: 0200049  
 Column: 250\*4.6mm, Kromasil-C18-5um  
 Solvent A: 0.1%TFA in 100%water  
 Solvent B: 0.1%TFA in 100%acetonitrile  
 Gradient:
 

	A	B
0.1min	30%	70%
25.0min	55%	45%
25.1min	0%	100%
30.0min	stop	

Flow Rate: 1.0ml/min  
 Wavelength(nm): 220  
 Volume: 5ul



Peak No.	Ret Time	Height	Area	Conc.
1	11.968	1435.236	6138.164	0.3083
2	12.258	193891.516	1963054.625	98.5942
3	12.258	2921.095	9570.748	0.4807
4	12.823	1113.781	12281.146	0.6168